

CERTIFICATE OF TYPE TESTING IN ACCORDANCE WITH BS EN 14175

CERTIFICATE & REPORT NO: INV/EN14175/422

DATE: 17th July 2008



Fume Cupboard Manufacturer:
Chemical Systems Control Ltd
116 Ashbourne Industrial Estate
Ashbourne, Co Meath
IRELAND

Fume Cupboard Type:
ChemFlow - 1.5m wide bench-type
Test opening width: 1180mm
Test opening height: 500mm
Extract volume flow rate: 0.330m³/s
Average face velocity: 0.509m/s

This is to certify that the fume cupboard described above has been type-tested in accordance with Part 3, in compliance with the requirements of Part 2 and with reference to Part 1 of BS EN 14175 and resulted in the performance characteristics given in the corresponding test report.

Tested and Certified by: Dr A F Bicen
Invent UK Ltd, 85 Southdown Road, Harpenden, Herts AL5 1PR
Tel: 01582 461666 : Fax: 01582 460077 : E-mail: info@invent-uk.com

A handwritten signature in blue ink that reads 'A.F. Bicen'.